



University of
Salford
MANCHESTER

A systematic review of the effect of footwear, foot orthoses and taping on lower limb muscle activity during walking and running

Reeves, JE, Jones, R, Liu, A, Bent, L, Plater, E and Nester, CJ
<http://dx.doi.org/10.1177/0309364619870666>

Title	A systematic review of the effect of footwear, foot orthoses and taping on lower limb muscle activity during walking and running
Authors	Reeves, JE, Jones, R, Liu, A, Bent, L, Plater, E and Nester, CJ
Type	Article
URL	This version is available at: http://usir.salford.ac.uk/id/eprint/52632/
Published Date	2019

USIR is a digital collection of the research output of the University of Salford. Where copyright permits, full text material held in the repository is made freely available online and can be read, downloaded and copied for non-commercial private study or research purposes. Please check the manuscript for any further copyright restrictions.

For more information, including our policy and submission procedure, please contact the Repository Team at: usir@salford.ac.uk.

Abstract

Background:

External devices are used to manage musculoskeletal pathologies by altering loading of the foot, which could result in altered muscle activity that could have therapeutic benefits.

Objectives:

To establish if evidence exists that footwear, foot orthoses and taping alter lower limb muscle activity during walking and running.

Study design:

Systematic literature review.

Methods:

CINAHL, MEDLINE, ScienceDirect, SPORTDiscus and Web of Science databases were searched. Quality assessment was performed using guidelines for assessing healthcare interventions and electromyography methodology.

Results:

Thirty-one studies were included: 22 related to footwear, eight foot orthoses and one taping. In walking: 1) Rocker footwear apparently decreases tibialis anterior activity and increases triceps surae activity; 2) Orthoses could decrease activity of tibialis posterior and increase activity of peroneus longus; 3) Other footwear and taping effects are unclear.

Conclusion:

Modifications in shoe or orthosis design in the sagittal or frontal plane can alter activation in walking of muscles acting primarily in these planes. Adequately powered research with kinematic and kinetic data is needed to explain the presence/absence of changes in muscle activation with external devices.

Clinical relevance:

This review provides some evidence that foot orthoses can reduce tibialis posterior activity, potentially benefiting specific musculoskeletal pathologies.

Keywords:

Rocker footwear; foot orthoses; tibialis posterior; electromyography

Word count:
4,981

Background

Musculoskeletal pathologies occur when structures experience more load than they can withstand.¹ If external loads are altered by a therapeutic device there should be a corresponding change in internal muscle-tendon forces, joint loading, the potential for injury, and the rate and likelihood of healing. For example, foot orthoses (FOs) that decrease loading at the rearfoot could decrease activity of the tibialis posterior (TP) muscle and subsequently reduce strain of the TP tendon, a structure vulnerable to tendinopathy.² Clinicians can influence the forces applied to feet and muscles/tendons using footwear, FOs and taping.

Footwear that may have therapeutic benefits by altering loading of the foot include ‘motion control’ shoes, (typically running shoes), and rocker/rollover shoes. Whilst motion control shoes with dual density midsoles reduce calcaneal eversion by 2.77° ($p < 0.001$, 95% CI 1.74° to 3.81°),³ whether these changes impact on muscle function and injury risk is unknown. Under the “preferred movement pathway” theory,⁴ footwear or FOs reduce muscle activity and metabolic demand^{4, 5} by promoting the path of “least resistance” and reduce injury risk.⁴⁻⁶ However, muscle activity could also increase to keep foot kinematics within the preferred pathway.⁶ Rocker or rollover shoes have outsoles curved in the sagittal plane and alter the contact area between the shoe and floor, plantar load, external sagittal plane joint moments, and thereafter muscular responses and joint motion.⁷ A recent review, however, found few statistically significant effects of the Masai Barefoot Technology (MBT) shoes on lower limb muscles.⁸ The effect of other rocker and motion control footwear on EMG data has not been reviewed, so the use of these specialised shoes for treatment and injury prevention is unclear.

Foot orthoses redistribute plantar pressure, altering external joint moments, internal joint moments (from muscles and connective tissue), and foot motion. Examples include insoles with rearfoot wedges and arch supports,⁹ a.k.a. “anti-pronation” FOs or medial posted FOs. Although FOs reduce peak rearfoot eversion by 2.08° to 2.35° ($p \leq 0.004$) depending on their design,³ such small changes may not be clinically meaningful.¹⁰ Foot orthoses can change ankle moments¹¹⁻¹³ with peak and mean ankle eversion moments reduced by $1.1 \pm 1.1\%$ ($p = 0.003$) and $2.3 \pm 2.1\%$ ($p < 0.001$) per 2° of medial posting respectively.¹³ Such changes would alter the requirements of tissues acting antagonistically to the external moments, including muscles. The evidence for changes in muscle function with FOs that alter joint kinematics and kinetics is important in understanding injury risk and tissue repair.

Low-Dye taping is a temporary intervention for conditions supposedly associated with foot pronation or flat-arched feet.^{14, 15} Theoretically applying tension to the skin using tape offloads

structures in the medial arch.^{3, 14, 16} Taping has been reported to only reduce foot pronation by a non-significant 1.50° (p= 0.19, 95% CI=-0.73° to 3.73°).³ Plantar sensory stimulation is considered an important difference between FOs and taping since changes in afferent feedback due to tape might alter muscle activation.¹⁴ However, there is no evidence to support this theory.

Prior reviews investigating the effects of footwear, FOs and taping did not compare device effects.^{8, 16, 17} Also, approaches to searching and appraisal of literature was variable and underpin the need for a more comprehensive review. Indeed limitations of prior studies include low power, inadequate reporting of electromyography (EMG) procedures and low external validity.^{16, 17} A review of foot posture, FOs and footwear by Murley et al.¹⁷ allowed comparisons to a barefoot control, which is less clinically generalizable than a shod control and was broad, including all types of FOs and inserts and all footwear, not just that intended to alter foot biomechanics. The present review includes only FOs with a medial arch profile and or medial heel/foot wedge in order to improve our understanding of the relationship between medial FOs and muscle activity. Furthermore the recent review of MBT footwear did not assess the quality of EMG data reporting, limiting our understanding of the strength of the evidence identified.⁸ Several studies on the effect of external devices on EMG have been published since these reviews, some of which have reported detailed EMG methods,¹⁸ thus further justifying an update on the literature consensus. The aim of this systematic review was to investigate the level of evidence from any study design that investigated whether footwear, FOs and taping alter lower limb EMG during walking and running, irrespective of health status.

Methods

Search strategy

A systematic, electronic database search was performed by reviewer J.R. using CINAHL (1982-2017), MEDLINE (1950-2017), ScienceDirect, SPORTDiscus (1985-2017) and Web of Science (1900-2017) in October 2015 and updated in March 2019. The review conformed to the PRISMA guidelines for systematic reviews, however we were unable to account for biases like publication bias.¹⁹ Searched words are included in Table 1. Lines 1-3) were combined using “AND” with lines 4) and 5). Additional sources were identified from published reviews and the reference lists of studies that passed the quality screening.

Inclusion criteria

The search results were assessed for eligibility based on titles and abstracts of original, full text articles using the following inclusion criteria:

- 1) A clearly defined amplitude, timing or frequency EMG outcome measure from muscles of the lower limb.
- 2) A fully specified independent variable of any footwear designed with modifications in the shape or material of the sole (including a negative heel, but excluding high heels, ankle braces and ankle destabilisation devices), foot orthoses/insoles (orthosis had a medial arch profile and or medial heel/foot wedge, excluding lateral wedges and ankle-foot orthoses), and taping about the foot/ankle intended to reduce foot pronation (excluding Kinesio taping).
- 3) Measures were made during level walking or running.
- 4) The footwear, FOs or taping experimental conditions were compared with a shod control condition.
- 5) For FOs and taping experimental conditions trials were performed in shoes, not sandals, with all the standard components of a shoe that brace the FOs.
- 6) Participants were free from conditions affecting the neurological systems.
- 7) Data was analysed from a minimum of three trials per condition.
- 8) Full text was published in English, French or German (due to available expertise).
- 9) Sample size of $n > 1$.

Only studies on locomotion were included since major theories on mechanisms of therapeutic effect of external devices relate to gait not standing.²⁰ Studies that only compared the device to barefoot were excluded because EMG amplitude can increase due to shoes alone and FOs versus barefoot (+30% and +30-38% respectively in tibialis anterior (TA)).²¹ Articles were excluded if there were less than three trials per condition because without contradictory evidence, this was considered the minimum required for quality data. We did not restrict studies to a specific population as we took a mechanistic approach to understanding potential effects of external devices on muscle activity.

Quality assessment

To maintain quality standards in this systematic review the articles that met the inclusion criteria were subject to two levels of quality assessment (Table 2), performed independently by reviewers J.R. and E.P. After studies were assessed the two reviewers met to discuss

discrepancies. When discrepancies persisted these were discussed with a third reviewer (L.B.) and a final score obtained.

The first stage focussed on the quality of the EMG methodology based on external standards of reporting,²² plus controlling locomotion velocity (since velocity can affect EMG).²³ Studies scored a 1 or 0 depending on whether the criteria was fulfilled or not and the results were summated and expressed as a percentage. Studies achieving less than 50% were excluded.

The second stage of assessment was based on a modified sub set of a checklist for rating clinical interventions.²⁴ Studies were given 1 or 0 depending on whether each criteria was fulfilled, with the total score expressed as a percentage and studies that scored less than 50% were excluded.

Results

Search results

A flow chart of the selection process from identification to screening and eligibility and inclusion¹⁹ is presented in Figure 1, six studies were excluded based on EMG quality and four studies were excluded based on study design quality. A total of 31 studies were included, of these, 22 related to footwear and eight to FOs. Key themes of footwear studies were running shoes, rocker footwear, APOS-Therapy shoes^{25, 26} and the Reebok EasyTone® shoe. Only one taping study (low-Dye) was identified which passed quality assessment. No study from additional sources met the inclusion criteria. Two studies were translated from German, but did not meet the inclusion criteria. Most studies included healthy, often recreationally active participants, except two studies involving participants with knee pain/knee osteoarthritis,^{27, 28} one with running related overuse injuries²⁹ and one with Achilles tendinopathy.³⁰ Summaries of included studies are in Tables 3-4 and excluded studies in the appendix.

Quality assessment

The included studies scored 50-100 on EMG quality (mean \pm SD: 75 \pm 13). Unfulfilled criteria for EMG quality spanned categories 1-6. Almost all studies were deemed to have provided adequate details on normalisation where appropriate, although none as detailed as published recommendations e.g. training to produce a maximum voluntary contraction (MVC).²² Of included studies 14/31 did not specify EMG sensor fixation and 13/31 did not control velocity.

Based on the second stage of quality assessment, the included studies were generally of moderate quality (scored 54-85, mean \pm SD: 64 \pm 9). Excluded studies scored 31-46 (mean \pm SD: 40 \pm 7). Many studies did not report device material, participant and assessor blinding, statistical analysis or power analysis. Variability was not reported in 3/4 of excluded studies. Additionally condition randomisation was absent in 8/31 of the included studies and 2/4 of the excluded studies.

22, 31

Overview of included studies

Footwear

Running shoes

A stability running shoe with a dual-density medial post and foot bridge had no effect on EMG activity during walking versus a standard flexible shoe for the peroneus longus (PL), medial gastrocnemius (MG), soleus, tibialis anterior (TA) and TP.³² However a motion control running shoe with a dual-density midsole (firmer material on the medial versus lateral side), reduced mean TA and PL activity and delayed fatigue during running³³ versus a cushioned running shoe. Delayed fatigue was demonstrated by maintained median frequency of TA and PL during a 10 km run.

A shoe with a medial wedge (3 cm thick on the medial side, 2 cm thick on the lateral side) increased mean TA amplitude by 16% during treadmill running versus a neutral shoe,³⁴ but not during overground running.³⁵ The activity of the gastrocnemii, soleus, PL and TP were not affected by medial or lateral wedges.^{34, 35} However TP EMG data was only available for four participants due to measurement difficulties.^{34, 35}

The relative timing of vastus medialis (VM) and vastus lateralis (VL) activation during running was compared between a motion control running shoe with a dual-density midsole and a cushioning running shoe.³⁶ This comparison was made based on the premise that delayed onset of VM with respect to VL is associated with patellofemoral pain.³⁶ The authors normalised EMG signals to a “duty cycle” (defined in the animal literature as stance expressed as a percentage of step cycle, i.e. stance + swing).³⁷ In the motion control shoe activation of VM occurred ~5.3%

7, 27, 30, 32-36, 38-52

(95% CI 4.5% to 6.1%) of a duty cycle earlier than VL (during a 10 km run). In contrast VM activation occurred ~4.6% (95% CI 3.9% to 5.3%) later than VL in the neutral shoe.³⁶ The implications of the findings are limited by the ambiguity of the reporting of the methods.

Rocker footwear

Tibialis anterior amplitude in early stance in walking reduced by ~30-40% ($p < 0.05$) with rocker footwear like MBT, which is curved under the heel in the sagittal plane, versus flat heeled conventional footwear.^{7, 50, 51} There was also a trend towards reduced TA EMG intensity when walking in MBT versus a running shoe.⁴⁸ A modified shoe with a forefoot rocker only *increased* peak TA activity in walking (by 20-35%, $p < 0.001$, $p = 0.015$ respectively), but not in running.^{30, 52} The shoe did not alter triceps surae (TS) activity in late stance during either walking or running.^{30, 52}

Increased PL activity throughout stance (e.g. 50% at loading response, $p = 0.02$) has been shown with MBT.⁵⁰ Other footwear in that study, including FitFlops™, designed to be unstable in the sagittal plane, increased peroneal activity during pre-swing. However, later work found no difference in PL activity with FitFlops™⁴⁹ or difference in co-contraction with regular flip-flops.⁴⁰ A minimalist shoe reduced TA activity and increased plantar flexion in early stance relative to control footwear, however walking speed was also slower.⁴²

Changes in TS activity during loading response with rocker footwear are opposite to that of TA.^{7, 50} The integral of the EMG profile in rocker footwear was 8-13% ($p < 0.05$) greater than the control shoe for the soleus and 5.5-8% for the MG (significant for MBT, $p < 0.05$, but not other rocker shoe, $p > 0.05$).⁷ Similarly, integrated EMG of MG was 6-16% ($p < 0.05$) higher and 8-23% ($p < 0.01$) higher in soleus in a rocker shoe compared with a regular walking shoe in treadmill walking.⁴⁷ There was also a trend towards increased MG EMG intensity walking in MBT compared with a running shoe in another study.⁴⁸ Activation of MG was unaffected when wearing a FitFlop™ sandal with a variable density sole.^{39, 50}

Studies recording quadriceps activation during walking have generally found no effect of rocker footwear.^{7, 48, 50, 51} However increased activation of VL and greater co-contraction of vastii and MG across stance was found in MBT.^{38, 45} Activation of biceps femoris or rectus femoris was unaltered by a FitFlop™ sandal.³⁹ Stiff soled safety shoes significantly increased VL, biceps femoris and TA activity relative to a soft soled trainer.⁴³

APOS-Therapy shoes

APOS-Therapy shoes have adjustable domes on the sole allowing manipulation of COP position and external joint moments.^{25, 26} In females with knee osteoarthritis a lateral shift in the sole domes and COP reduced averaged TA EMG amplitude in pre-swing versus a neutral dome configuration.²⁷ The EMG amplitude of the lateral gastrocnemius increased with a medial shift in COP and decreased with a lateral shift in COP due to APOS-Therapy shoes (compared with neutral).²⁷

Reebok EasyTone® shoe

The Reebok EasyTone® shoe, designed to be unstable with balance pods, did not alter muscle activation in walking of any thigh, shank or gluteal muscles.^{41, 44, 45}

Foot orthoses

There is some limited evidence that FOs decrease activity of TP in early stance and increase activity of PL in mid-late stance.^{9, 21} Peak amplitude and RMS amplitude of TP during loading response was shown to reduce by 19% ($p=0.007$) and 22% ($p=0.002$) respectively with custom FOs, and 12% ($p<0.001$) and 13% ($p=0.001$) respectively with prefabricated FOs.⁹ Whereas PL activity increased in midstance with a prefabricated FOs (peak amplitude +21%, $p=0.024$; RMS amplitude +24%, $p=0.019$) and a custom FOs (peak amplitude +16%, $p=0.028$) compared with a shoe only. Maximum PL amplitude has also been shown to increase in walking by 19% for pronated individuals when wearing 15° inverted FOs versus shoes alone ($p<0.05$).²¹ However, PL amplitude does not appear to increase linearly with wedging magnitude.^{18, 21} Another two studies found TP activity was not significantly different between the footwear and FOs conditions ($p>0.05$), although there was a decrease of around 10% ($p<0.05$) from barefoot to shod and shod with either a prefabricated or custom FOs, which was not considered clinically generalizable in this review.^{2, 53} However in the study that recorded kinematics and kinetics, there was also no effect of FOs on subtalar joint displacement or supination moment relative to the shoe condition.² There was no difference in flexor digitorum longus or PL activity between conditions.⁵³

As for TA and TS, most evidence indicates magnitude and timing of activation is unchanged by wearing FOs during walking and running.^{18, 21, 28, 54} One study found there was a tendency for FOs to decrease TA activation during walking versus a shoe only (effect size 0.18-0.29, for custom and prefabricated FOs respectively), although the result was not statistically significant.⁹

Activity of PL may also increase with FOs during running. In one study 99 runners with an overuse injury were assigned to customised FOs or no FOs.²⁹ In treadmill running, there was a significant increase ($p= 0.003$) in PL pre-activation amplitude (EMG activity prior to foot contact) after two months wearing FOs, but not in the control patient group.²⁹ It is unclear whether change was in barefoot running or the running shoes or both. Another study reported a 14.7% ($p < 0.05$) greater duration of PL activity (the muscle was active for longer) during running with prefabricated FOs compared with no FOs as well as lower average MG and VM RMS amplitude with FOs versus no FOs.⁵⁴

Low-Dye taping

In the only study included involving taping a significant delay (5-7%, $p= 0.001$) in onset times of VM, VL and gluteus medius was found during shod running with low-Dye tape compared with control taping.⁴⁶

Discussion

The aim of this review was to establish if there is evidence that footwear, FOs and taping alter muscle activity of the lower limb during walking and running. The effect of running shoe design, FitFlops™ sandals and low-Dye taping on muscle activity is unclear, while rocker/rollover shoes appear to affect muscle activity of MG and TS.^{7, 47, 50} There is evidence, albeit limited, that FOs decrease activity of TP in early stance,⁹ which could be beneficial in treating posterior tibial tendon dysfunction (PTTD). Activity of PL may increase in mid-late stance,^{9, 21} otherwise FOs do not appear to alter EMG of lower limb muscles.^{9, 18, 21, 28}

Footwear

Running shoes

The effect of running shoe design on muscle activity remains unclear due to uncertainty regarding whether the shoes tested were effective in changing loading. No study investigating the effects of running shoe design on EMG during walking or running collected simultaneous kinematics or kinetics.^{32, 33, 36} Without kinetic data we cannot determine if the footwear changed loading of the foot, which might sometimes explain the absence of change in EMG. Concurrent collection of kinetic and EMG data would also be useful to establish if the difference between a nil effect of motion control shoes in walking and a reduction in fatigue during running are due to the greater forces in running, foot strike patterns, or different shoe properties.^{32, 33}

Similarly sagittal plane kinematics were not reported in studies involving a medial wedged shoe.^{34, 35} Wedging could increase ankle plantar flexion and increase demand on TA, potentially explaining the 16% increase in TA amplitude during treadmill running compared with a neutral shoe.³⁴ Perhaps a medial wedged shoe is substantially different to a motion control running shoe with a dual-density midsole if the effects of the wedge are not isolated to the frontal plane.

Rocker footwear

Footwear that shifts the COP anteriorly at heel contact reduces TA amplitude between initial contact and into midstance.^{7, 50, 51} An anterior shift in the COP increased the external dorsiflexion moment, resulting in a more dorsiflexed ankle and less work required from TA to control plantarflexion after initial contact.⁷ Increased external dorsiflexion moment/increased internal plantar flexion moment in early stance would also account for the increases in TS activity with some rocker footwear.^{7, 47, 50} Potentially the increased PL activity in MBT shoes⁵⁰ is due to the need for PL to contribute to sagittal plane moments. In contrast the *increase* in TA activity in walking with the shoe with the modified forefoot rocker^{30, 52} might be explained by the greater mass and sole thickness of the modified shoe versus the control. As TA is active in swing, greater activity during early stance could result from the greater moment of inertia not the sole curvature.

Rocker footwear have been shown to reduce internal plantar flexion moment in late stance, which could be beneficial for offloading the Achilles tendon when treating Achilles tendinopathy.^{8, 30, 52} However reduced internal plantar flexion moment in late stance is not necessarily coupled to reduced TS activity in the same phase.^{30, 50, 52} This could be because peak o MG activity is earlier in stance than the peak of the internal plantar flexion moment⁷ and the energy recoil of the Achilles tendon is in terminal stance. Thus reduced loading of the Achilles tendon suggested by a reduced internal plantar flexion moment may still be beneficial in treating Achilles tendinopathy.

The curved sole of rocker footwear purportedly reduces contact area with the ground and thus reduces stability. Increased co-activation from TA and TS in early stance with MBT may increase ankle stability to compensate.⁷ The induced instability is assumed to increase movement variability and activate muscles required to maintain balance and control movement.⁵⁵ Greater movement variability could be beneficial in managing chronic injury if it reduced the repetitive loading of injured structures.⁵⁶ Conversely increased co-activation increases joint loading.⁵⁷ Consequently the clinical implications of altered muscle activation

and reduced internal plantar flexion with rocker footwear remains unestablished. Additionally there was no effect of the modified rocker shoe on pain in individuals with chronic Achilles tendinopathy and randomised clinical trials are necessary to evaluate the therapeutic effect of rocker footwear.³⁰

APOS-Therapy shoes

Reduced TA and increased TS activation with APOS-Therapy shoes could be relevant to those with TS and Achilles injury, anterior compartment syndrome and intermittent claudication, but any implications remain speculative.

Foot orthoses

There is some limited evidence that FOs decrease activity of TP in early stance and increase activity of PL in mid-late stance,^{9, 21} but otherwise there appears to be a lack of effect of FOs on lower limb muscle activity during walking.^{9, 18, 21, 28, 54}

As the primary inverter of the foot tibialis posterior (TP) acts eccentrically during early stance to generate an inversion moment that opposes the external eversion moment, and helps control rearfoot eversion. It also acts concentrically to support foot supination later in stance.⁵⁸ If FOs increase the external inversion moment, they might reduce required internal inversion moments, reducing TP activity. Reduced TP activation could mean less force through the TP tendon which could facilitate healing in pathologies like PTTD. The limited amount of evidence on the effect of FOs on TP is likely because indwelling EMG is required to measure TP activity. Further research with adequate power and concurrent collection of kinematic and kinetic data is needed to relate kinetic and kinematic changes to muscle activation.

A linear dose-response to extrinsic rearfoot posting during walking has been demonstrated in kinematic, kinetic and plantar pressure variables, but without a corresponding effect on any EMG related muscle activity in the calf muscles (including PL), quadriceps or hamstrings.^{13, 18} Maximum PL amplitude did increase in walking by 19% for pronated individuals when wearing 15° inverted FOs versus shoes alone, but again without a linear dose-response to magnitude of wedging.²¹ The lack of a dose response to medial rearfoot wedging could infer that the FOs exert their effect on PL due to changes in load under the medial longitudinal arch rather than the rearfoot.²¹ The midfoot is in contact with the ground during midstance and the heel is unloading.⁵⁹ Similarly in later work by Murley and colleagues, flat-footed participants increased PL activity in midstance with prefabricated FOs (peak amplitude +21%, $p= 0.024$;

RMS amplitude +24%, $p=0.019$) and custom FOs (peak amplitude +16%, $p=0.028$) compared with a shoe only.⁹ The original authors speculated that increased PL EMG amplitude resulted from the foot being more laterally unstable. If FOs increased the external inversion moment, greater PL EMG activation may be needed to maintain equilibrium. As PL is the antagonist of TP, if FOs reduced TP EMG activity this would possibly be accompanied by increased PL EMG activity. However, TP and PL activity do not necessarily represent equal opposing inversion and eversion moments respectively, due to additional muscle tendon parameters⁶⁰ and different moment arms.

Although FOs may increase amplitude and duration of PL activity during running,^{29, 54} the literature is limited by low between-session reliability of PL EMG.⁶¹⁻⁶³ Reported poor inter-session reliability of EMG data from PL reduces confidence in EMG results collected weeks apart.⁶¹⁻⁶³ Amplitude of an EMG signal varies not only due to the detection of different motor units, but because of variable skin-electrode impedance between sessions.⁶⁴ Variability in amplitude between sessions could affect the ability to detect changes in duration of muscle activity due to FOs using threshold methods. As measurements were taken in separate sessions without mention of normalisation^{29, 54} comparing EMG measures could be beyond this technique. Additionally electrode placement in one study²⁹ followed the methods of Winter and Yack⁶⁵ (50% of the distance between the fibular head and lateral malleolus, rather than 25% of the distance recommended by SENIAM).³¹ A distal shift in surface electrode placement over PL of 2 cm increases the presence of crosstalk, likely from TA.⁶⁶ Given that PL is most active in mid-late stance, while TA is active prior to foot contact to enable a dorsiflexed ankle position at initial contact, potentially muscle activity reported as pre-activation of PL was actually crosstalk from TA.

Foot orthoses designed to reduce the external eversion moment at the subtalar joint would theoretically decrease the internal inversion moment required from the invertor muscles limiting eversion. The TA has an inversion moment arm when the foot is inverted, as at initial contact,⁶⁷ therefore FOs that reduce the eversion moment might also reduce TA activity. The conclusion of the review by Murley et al.¹⁷ preceding the work of Telfer^{13, 18} that FOs may *increase* activation of the TA should be reconsidered. Studies that found FOs increase TA activity had notable limitations. As Murley et al.¹⁷ identified, a significant increase in EMG activation was not always supported by confidence intervals. Many studies did not simultaneously collect kinematic and kinetic data so we cannot relate any change (or lack of) in EMG to other changes in biomechanics, or evaluate the intervention in the context of the

“preferred movement pathway” theory. For example, an extrinsic medial rearfoot wedge could place the foot in a more plantar flexed position since the heel is lifted in the shoe, perhaps increasing demand on the TA in the sagittal plane after initial contact.³⁴ Without kinematic and kinetic data and with variable changes in EMG, the implications of this finding are limited. As TA is not the principal invertor of the foot and its main role is dorsiflexion, perhaps any effect is too small to detect, or too variable depending upon the action of the other invertor muscles (i.e. posterior leg muscles passing medial to the ankle) and foot position. Also, as the only ankle dorsiflexor, TA function is unlikely to be compromised with more alternative invertor muscles available. Overall the majority of studies have found FOs do not change TA activity significantly, in some cases FOs may decrease TA activity, but any effect is subtle.

Literature limitations

Tibialis posterior is the largest invertor, but given fine-wire EMG can be challenging few studies have investigated its function, or intrinsic muscle activity. Furthermore, the magnitude of change in muscle activity that is clinically meaningful is unknown, thus significant effects of external devices on EMG does not reveal clinically beneficial effects. Electromyography is only a measure of electrical activity not force production, nor mechanical work in the muscle-tendon unit. Additionally, differences in electrode types, signal processing, normalisation and outcome variables make establishing a consensus regarding the meaning of changes in EMG difficult. Guidelines describe methods of EMG processing, but are not a universal best practice.²²

Research investigating FOs used various materials and designs, and the descriptions of FOs were limited (no excluded study provided detail on this criteria). Studies used a mixture of customised and prefabricated FOs and both FOs with modifications only in the rearfoot and FOs with additional modifications in the arch and forefoot. Whether isolated modifications in specific FOs geometry could lead to specific changes in EMG is unclear. Additionally several studies may have also been inadequately powered. A final observation is that studies generally focus on the immediate effect of external devices, yet muscle function could change over time. Longitudinal EMG studies are difficult, but other approaches such as muscle morphology have proven sensitive to footwear.^{68, 69}

Review limitations

While the quality assessment allowed the review to be based on studies of at least moderate quality, failing criteria could reflect inadequacy in reporting and not whether appropriate

procedures were followed.¹⁷ Additionally, database searches from one reviewer and the score of 50% as a threshold for inclusion could be considered subjective. Furthermore each criteria was given equal weighting when some could be more influential than others. For instance blinding might be unrealistic, as so-called sham FOs can exert mechanical effects⁷⁰ and potentially influence EMG. Conversely, lack of randomisation could invalidate results due to an order effect and be grounds for exclusion alone. Nonetheless, outcomes of the excluded studies were largely in agreement with those included, except one which found a significantly longer duration ($p < 0.05$) of TA activity following foot contact with FOs versus control.⁷¹

Footwear outside the inclusion criteria could alter loading of the foot and subsequent muscle activity. However a general review of footwear would be far broader and by restricting our search to footwear that aims to alter foot/ankle motion with modifications in sole construction, findings can be more directly related to the other devices reviewed. This review focused on muscle activation, however devices could have other effects on soft tissue, like the capacity of the series elastic element of TP to absorb energy in early stance.⁵⁸

The review included studies with heterogeneous injury status and foot postures. The response to an intervention may vary with pathology. However the evidence that Achilles tendinopathy for example alters muscle activation is conflicting.^{72, 73} Few of the studies included patient populations and those that did did not provide healthy controls and foot posture was often not reported. Consequently sub-group comparisons were not possible.

Conclusion

Modifications in shoe or FOs design in the sagittal or frontal plane can alter activation in walking of muscles acting primarily in these planes. Adequately powered research with kinematic and kinetic data is needed to explain the presence/absence of changes in muscle activation with external devices.

Acknowledgements

This research received no specific grant from any funding agency in the public, commercial, or not-for-profit sectors. We wish to thank Dr. Henrike Greaves (Liverpool John Moores University) for her assistance in translating articles to English from German.

Declaration of conflicting of interests

C.N. owns equity in a company that manufactures foot orthoses. Other Authors have no conflicts of interest to declare.

The review was conducted by J.R. The preparation of the manuscript was supervised by C.N. All authors were involved in the drafting and approving of the manuscript.

References

1. Payne C. Forces, motion and outcomes with foot orthoses and running shoes. In: E. C. Frederick and S. W. Yang, (eds.). *8th Footwear Biomechanics Symposium*. Taipei, Taiwan 2007, p. 23-24.
2. Maharaj JN, Cresswell AG and Lichtwark GA. The Immediate Effect of Foot Orthoses on Subtalar Joint Mechanics and Energetics. *Med Sci Sports Exerc* 2018; 50: 1449-1456.
3. Cheung RTH, Chung RCK and Ng GYF. Efficacies of different external controls for excessive foot pronation: a meta-analysis. *Br J Sports Med* 2011; 45: 743-751.
4. Nigg BM. The role of impact forces and foot pronation: A new paradigm. *Clin J Sport Med* 2001; 11: 2-9.
5. Nigg BM, Vienneau J, Smith AC, et al. The Preferred Movement Path Paradigm: Influence of Running Shoes on Joint Movement. *Med Sci Sports Exerc* 2017; 49: 1641-1648.
6. Nigg B, Mohr M and Nigg SR. Muscle tuning and preferred movement path-a paradigm shift. *Current Issues in Sport Science (CISS)* 2017.
7. Forghany S, Nester CJ, Richards B, et al. Rollover footwear affects lower limb biomechanics during walking. *Gait Posture* 2014; 39: 205-212.
8. Tan JM, Auhl M, Menz HB, et al. The effect of Masai Barefoot Technology (MBT) footwear on lower limb biomechanics: A systematic review. *Gait Posture* 2016; 43: 76-86.
9. Murley GS, Landorf KB and Menz HB. Do foot orthoses change lower limb muscle activity in flat-arched feet towards a pattern observed in normal-arched feet? *Clin Biomech* 2010; 25: 728-736.
10. Chevalier TL and Chockalingam N. Foot Orthoses A Review Focusing on Kinematics. *J Am Podiatr Med Assoc* 2011; 101: 341-348.
11. McMillan A and Payne C. Effect of foot orthoses on lower extremity kinetics during running: a systematic literature review. *J Foot Ankle Res* 2008; 1: 13.
12. Nester CJ, van der Linden ML and Bowker P. Effect of foot orthoses on the kinematics and kinetics of normal walking gait. *Gait Posture* 2003; 17: 180-187.
13. Telfer S, Abbott M, Steultjens MPM, et al. Dose response effects of customised foot orthoses on lower limb kinematics and kinetics in pronated foot type. *J Biomech* 2013; 46: 1489-1495.

14. Bishop C, Arnold JB and May T. Effects of Taping and Orthoses on Foot Biomechanics in Adults with Flat-Arched Feet. *Med Sci Sports Exerc* 2016; 48: 689-696.
15. Radford JA, Burns J, Buchbinder R, et al. The effect of low-dye taping on kinematic, kinetic, and electromyographic variables: A systematic review. *J Orthop Sports Phys Ther* 2006; 36: 232-241.
16. Franettovich M, Chapman A, Blanch P, et al. A physiological and psychological basis for anti-pronation taping from a critical review of the literature. *Sports Med* 2008; 38: 617-631.
17. Murley GS, Landorf KB, Menz HB, et al. Effect of foot posture, foot orthoses and footwear on lower limb muscle activity during walking and running: a systematic review. *Gait Posture* 2009; 29: 172-187.
18. Telfer S, Abbott M, Steultjens M, et al. Dose–response effects of customised foot orthoses on lower limb muscle activity and plantar pressures in pronated foot type. *Gait Posture* 2013; 38: 443-449.
19. Liberati A, Altman DG, Tetzlaff J, et al. The PRISMA statement for reporting systematic reviews and meta-analyses of studies that evaluate health care interventions: explanation and elaboration. *Ann Intern Med* 2009; 151: W-65-W-94.
20. Harradine P and Bevan L. A Review of the Theoretical Unified Approach to Podiatric Biomechanics in Relation to Foot Orthoses Therapy. *J Am Podiatr Med Assoc* 2009; 99: 317-325.
21. Murley GS and Bird AR. The effect of three levels of foot orthotic wedging on the surface electromyographic activity of selected lower limb muscles during gait. *Clin Biomech* 2006; 21: 1074-1080.
22. Merletti R and Di Torino P. Standards for reporting EMG data. *J Electromyogr Kinesiol* 1999; 9: 3-4.
23. Murley GS, Menz HB and Landorf KB. Electromyographic patterns of tibialis posterior and related muscles when walking at different speeds. *Gait Posture* 2014; 39: 1080-1085.
24. Downs SH and Black N. The feasibility of creating a checklist for the assessment of the methodological quality both of randomised and non-randomised studies of health care interventions. *J Epidemiol Community Health* 1998; 52: 377-384.
25. Haim A, Rozen N, Dekel S, et al. Control of knee coronal plane moment via modulation of center of pressure: A prospective gait analysis study. *J Biomech* 2008; 41: 3010-3016.

26. Haim A, Rozen N and Wolf A. The influence of sagittal center of pressure offset on gait kinematics and kinetics. *J Biomech* 2010; 43: 969-977.
27. Goryachev Y, Debbi EM, Haim A, et al. Foot center of pressure manipulation and gait therapy influence lower limb muscle activation in patients with osteoarthritis of the knee. *J Electromyogr Kinesiol* 2011; 21: 704-711.
28. Mills K, Blanch P and Vicenzino B. Comfort and midfoot mobility rather than orthosis hardness or contouring influence their immediate effects on lower limb function in patients with anterior knee pain. *Clin Biomech* 2012; 27: 202-208.
29. Baur H, Hirschmuller A, Muller S, et al. Neuromuscular Activity of the Peroneal Muscle after Foot Orthoses Therapy in Runners. *Med Sci Sports Exerc* 2011; 43: 1500-1506.
30. Sobhani S, Zwerver J, van den Hetivel E, et al. Rocker shoes reduce Achilles tendon load in running and walking in patients with chronic Achilles tendinopathy. *Journal of Science and Medicine in Sport* 2015; 18: 133-138.
31. Hermens HJ, Freriks B, Disselhorst-Klug C, et al. Development of recommendations for SEMG sensors and sensor placement procedures. *J Electromyogr Kinesiol* 2000; 10: 361-374.
32. Scott LA, Murley GS and Wickham JB. The influence of footwear on the electromyographic activity of selected lower limb muscles during walking. *J Electromyogr Kinesiol* 2012; 22: 1010-1016.
33. Cheung RTH and Ng GYF. Motion Control Shoe Delays Fatigue of Shank Muscles in Runners With Overpronating Feet. *Am J Sports Med* 2010; 38: 486-491.
34. O'Connor KM, Price TB and Hamill J. Examination of extrinsic foot muscles during running using mfMRI and EMG. *J Electromyogr Kinesiol* 2006; 16: 522-530.
35. O'Connor KM and Hamill J. The role of selected extrinsic foot muscles during running. *Clin Biomech* 2004; 19: 71-77.
36. Cheung RTH and Ng GYF. Motion control shoe affects temporal activity of quadriceps in runners. *Br J Sports Med* 2009; 43: 943-947.
37. Liu Y, Ao LJ, Lu G, et al. Quantitative gait analysis of long-term locomotion deficits in classical unilateral striatal intracerebral hemorrhage rat model. *Behav Brain Res* 2013; 257: 166-177.
38. Buchecker M, Wagner H, Pfusterschmied J, et al. Lower extremity joint loading during level walking with Masai barefoot technology shoes in overweight males. *Scand J Med Sci Sports* 2012; 22: 372-380.

39. Burgess K and Swinton P. Do Fitflops™ increase lower limb muscle activity? *Clin Biomech* 2012; 27: 1078-1082.
40. Chen TL-W, Wong DW-C, Xu Z, et al. Lower limb muscle co-contraction and joint loading of flip-flops walking in male wearers. *PLoS One* 2018; 13.
41. Elkjaer EF, Kromann A, Larsen B, et al. EMG Analysis of Level and Incline Walking in Reebok EasyTone ET Calibrator. In: Dremstrup K, Rees S and Jensen O (eds) *15th Nordic-Baltic Conference on Biomedical Engineering and Medical Physics*. 2011, pp.109-112.
42. Franklin S, Li F-X and Grey MJ. Modifications in lower leg muscle activation when walking barefoot or in minimalist shoes across different age-groups. *Gait Posture* 2018; 60: 1-5.
43. Goto K and Abe K. Gait characteristics in women's safety shoes. *Appl Ergon* 2017; 65: 163-167.
44. Horsak B and Baca A. Effects of toning shoes on lower extremity gait biomechanics. *Clin Biomech* 2013; 28: 344-349.
45. Horsak B, Heller M and Baca A. Muscle co-contraction around the knee when walking with unstable shoes. *J Electromyogr Kinesiol* 2015; 25: 175-181.
46. Kelly LA, Racinais S, Tanner CM, et al. Augmented Low Dye Taping Changes Muscle Activation Patterns and Plantar Pressure During Treadmill Running. *J Orthop Sports Phys Ther* 2010; 40: 648-655.
47. Koyama K, Naito H, Ozaki H, et al. Effects of unstable shoes on energy cost during treadmill walking at various speeds. *J Sports Sci Med* 2012; 11: 632-637.
48. Nigg B, Hintzen S and Ferber R. Effect of an unstable shoe construction on lower extremity gait characteristics. *Clin Biomech* 2006; 21: 82-88.
49. Price C, Andrejevas V, Findlow AH, et al. Does flip-flop style footwear modify ankle biomechanics and foot loading patterns? *J Foot Ankle Res* 2014; 7.
50. Price C, Smith L, Graham-Smith P, et al. The effect of unstable sandals on instability in gait in healthy female subjects. *Gait Posture* 2013; 38: 410-415.
51. Sacco ICN, Sartor CD, Cacciari LP, et al. Effect of a rocker non-heeled shoe on EMG and ground reaction forces during gait without previous training. *Gait Posture* 2012; 36: 312-315.
52. Sobhani S, Hijmans J, van den Heuvel E, et al. Biomechanics of slow running and walking with a rocker shoe. *Gait Posture* 2013; 38: 998-1004.

53. Akuzawa H, Imai A, Iizuka S, et al. Calf muscle activity alteration with foot orthoses insertion during walking measured by fine-wire electromyography. *Journal of physical therapy science* 2016; 28: 3458-3462. 2017/02/09.
54. Kelly LA, Girard O and Racinais S. Effect of Orthoses on Changes in Neuromuscular Control and Aerobic Cost of a 1-h Run. *Med Sci Sports Exerc* 2011; 43: 2335-2343.
55. Nigg B, Federolf PA, von Tscharner V, et al. Unstable shoes: functional concepts and scientific evidence. *Footwear Sci* 2012; 4: 73-82.
56. Hamill J, van Emmerik RE, Heiderscheit BC, et al. A dynamical systems approach to lower extremity running injuries. *Clin Biomech* 1999; 14: 297-308.
57. Potthast W, Lersch C, Segesser B, et al. Intraarticular pressure distribution in the talocrural joint is related to lower leg muscle forces. *Clin Biomech* 2008; 23: 632-639.
58. Maharaj JN, Cresswell AG and Lichtwark GA. The mechanical function of the tibialis posterior muscle and its tendon during locomotion. *J Biomech* 2016; 49: 3238-3243.
59. Williams A and Nester C. CHAPTER 1 - Principles of foot biomechanics and gait. *Pocket Podiatry: Footwear and Foot Orthoses*. Edinburgh: Churchill Livingstone, 2010, pp.1-14.
60. Murley GS, Buldt AK, Trump PJ, et al. Tibialis posterior EMG activity during barefoot walking in people with neutral foot posture. *J Electromyogr Kinesiol* 2009; 19: E69-E77.
61. Barn R, Rafferty D, Turner DE, et al. Reliability study of tibialis posterior and selected leg muscle EMG and multi-segment foot kinematics in rheumatoid arthritis associated pes planovalgus. *Gait Posture* 2012; 36: 567-571.
62. Moisan G and Cantin V. Effects of two types of foot orthoses on lower limb muscle activity before and after a one-month period of wear. *Gait Posture* 2016; 46: 75-80.
63. Murley GS, Menz HB, Landorf KB, et al. Reliability of lower limb electromyography during overground walking: A comparison of maximal- and sub-maximal normalisation techniques. *J Biomech* 2010; 43: 749-756.
64. Burden A. Surface electromyography. In: Payton C and Bartlett R (eds) *Biomechanical evaluation of movement in sport and exercise: the British Association of Sport and Exercise Sciences guide*. Routledge, 2007, pp.77-102.
65. Winter D and Yack H. EMG profiles during normal human walking: stride-to-stride and inter-subject variability. *Electroencephalogr Clin Neurophysiol* 1987; 67: 402-411.

66. Campanini I, Merlo A, Degola P, et al. Effect of electrode location on EMG signal envelope in leg muscles during gait. *J Electromyogr Kinesiol* 2007; 17: 515-526.
67. Lee SSM and Piazza SJ. Inversion-eversion moment arms of gastrocnemius and tibialis anterior measured in vivo. *J Biomech* 2008; 41: 3366-3370.
68. Bruggemann G-P, Potthast W, Braunstein B, et al. Effect of increased mechanical stimuli on foot muscles functional capacity. In: *Proceedings of the ISB XXth Congress-ASB 29th Annual Meeting: 31 July-5 August 2005; Cleveland 2005*.
69. Johnson A, Myrer J, Mitchell U, et al. The effects of a transition to minimalist shoe running on intrinsic foot muscle size. *Int J Sports Med* 2016; 95: 154-158.
70. McCormick CJ, Bonanno DR and Landorf KB. The effect of customised and sham foot orthoses on plantar pressures. *J Foot Ankle Res* 2013; 6.
71. Tomaro J and Burdett RG. The effects of foot orthotics on the EMG activity of selected leg muscles during gait. *J Orthop Sports Phys Ther* 1993; 18: 532-536.
72. Baur H, Müller S, Hirschmüller A, et al. Comparison in lower leg neuromuscular activity between runners with unilateral mid-portion Achilles tendinopathy and healthy individuals. *J Electromyogr Kinesiol* 2011; 21: 499-505.
73. Wyndow N, Cowan S, Wrigley T, et al. Triceps surae activation is altered in male runners with Achilles tendinopathy. *J Electromyogr Kinesiol* 2013; 23: 166-172.